



JB36937

Customer: 1ST COMPUTER TRADERS
UNIT 6 MERCIAN PARK
FELSPAR ROAD
TAMWORTH
STAFFS
B77 4DT

Customer Order N°: 22746
Relevant Standards: BS EN ISO 14644
PD6609
BS EN ISO 14644-1 Classification: CLASS 5

Model N°:	K4	Serial No.	KV638/S41387
Commissioning / Revalidation Date:	19-Mar-07	Revalidation Due:	18-Mar-08
Commissioned / Revalidation By:	M JOHNSON	Witnessed By:	P ALLEN
Av. Filter Face Velocity:	0.45 m/s	Variac Setting:	70 %
Noise @ 1 Metre:	57 dB(A)	Filter Leak Test	Less than 0.001% detected
Max. Particle Count @ 0.5µm:	0		